

MEK-3DN HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B23YN
	2024-02-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	7.6	7.6	7.6	7.6	7.58
RBC	10 ⁶ /μL	±0.20	4.74	4.74	4.74	4.74	4.70
HGB	g/dL	±0.5	13.9	14.0	13.9	14.0	13.86
HCT	%	±3.0	40.0	40.0	40.0	40.0	45.5
MCV	fL	±7.0	84.4	84.4	84.4	84.4	96.8
MCH	pg	±2.4	29.3	29.5	29.3	29.5	29.5
MCHC	g/dL	±3.0	34.8	35.0	34.8	35.0	30.5
PLT	10 ³ /μL	±50	269	269	269	269	284.4
MPV	fL	±2.0	6.2	8.5	6.2	8.5	8.8
RDW-CV *	%	±3.0	17.1	17.1	17.1	17.1	16.7
LY%	%	±10.0	30.4	30.4	30.4	30.4	34.17
MO%	%	±5.0	4.4	4.4	4.4	4.4	2.54
GR%	%	±10.0	65.2	65.2	65.2	65.2	63.29
LY	10 ³ /μL	±1.0	2.3	2.3	2.3	2.3	2.59
MO	10 ³ /μL	±0.4	0.3	0.3	0.3	0.3	0.19
GR	10 ³ /μL	±1.0	5.0	5.0	5.0	5.0	4.80

MEK-1305
QC ESR
11±5 mm

Manual Registration	
Lot	23Y
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901685N